
Midmark M11 - v1.0.5
Total Cycles: 792

Name: RITTER2
Sterilizer ID: V2313228
02/14/2022 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.7 F 0.0
2:00 118.7 F 1.2
4:00 133.1 F 3.5
6:00 159.3 F 7.1
8:00 185.3 F 11.8
10:00 210.2 F 4.8
12:00 223.1 F 9.6
14:00 232.3 F 9.3
16:00 239.4 F 12.0
18:00 239.1 F 10.5
20:00 252.4 F 17.1
22:00 263.7 F 23.9
24:00 271.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.4 F 30.0
2:00 272.4 F 29.5
3:00 271.4 F 28.9
4:00 271.5 F 29.5

Min 271.2 F 28.9
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 221.9 F 0.0
10:00 229.0 F 0.0
15:00 222.9 F 0.0
20:00 217.6 F 0.0
25:00 217.5 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:07
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:02:09
CYCLE COMPLETE

02/14/2022 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____