
Midmark M11 - v1.0.0
Total Cycles: 4037

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2023 05:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.3 F 0.0
2:00 87.6 F 0.2
4:00 103.9 F 1.9
6:00 133.2 F 4.7
8:00 177.2 F 8.2
10:00 198.1 F 12.4
12:00 209.2 F 4.3
14:00 226.5 F 7.9
16:00 235.3 F 12.0
18:00 231.5 F 7.7
20:00 242.6 F 13.0
22:00 241.1 F 11.0
24:00 253.2 F 17.1
26:00 264.1 F 23.6
28:00 271.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.0
2:00 271.6 F 28.7
3:00 272.3 F 28.9
4:00 271.5 F 28.7

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 215.9 F 0.0
10:00 207.9 F 0.0
15:00 203.1 F 0.0
20:00 227.0 F 0.0
25:00 230.7 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 28:03
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:06:42
CYCLE COMPLETE

11/17/2023 07:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____