

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002474  
13:14 29/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:46  
STERILIZING 3:49  
136.5 C 225kPa 3:49  
136.4 C 225kPa 4:19  
136.4 C 225kPa 4:49  
136.3 C 222kPa 5:19  
136.4 C 224kPa 5:49  
136.2 C 223kPa 6:19  
136.2 C 222kPa 6:49  
136.4 C 225kPa 7:19  
136.4 C 225kPa 7:19  
VENTING 7:19  
AIR DRYING 8:09  
CYCLE COMPLETE 38:09

-----  
Digital Signature #  
7436EC69DB18A0CC  
-----