
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000896
13:37 02/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:22
STERILIZING 4:36
136.0 C 229kPa 4:36
136.4 C 230kPa 5:06
136.4 C 231kPa 5:36
136.4 C 233kPa 6:06
136.4 C 233kPa 6:36
136.4 C 232kPa 7:06
136.4 C 231kPa 7:36
136.0 C 229kPa 8:06
136.0 C 229kPa 8:06
VENTING 8:06
AIR DRYING 8:58
CYCLE COMPLETE 23:59

Digital Signature #
849075D9F8D1BFF7
