
Midmark M11 - v1.0.5
Total Cycles: 2288

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 82.5 F 0.0
4:00 126.6 F 0.0
6:00 116.5 F 0.1
8:00 150.3 F 2.3
10:00 183.1 F 5.3
12:00 214.8 F 9.0
14:00 224.7 F 13.6
16:00 227.9 F 7.6
18:00 238.2 F 11.9
20:00 233.2 F 8.6
22:00 240.9 F 11.0
24:00 244.5 F 13.0
26:00 256.6 F 19.4
28:00 267.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 271.8 F 29.0
2:00 272.1 F 29.3
3:00 272.2 F 29.6
4:00 271.7 F 29.5

Min 271.3 F 28.9
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 220.7 F 0.0
10:00 188.6 F 0.0
15:00 191.6 F 0.0
20:00 190.8 F 0.0
25:00 207.5 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:20
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:07:21
CYCLE COMPLETE

04/24/2023 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____