

-----  
Midmark M11 - v1.0.5  
Total Cycles: 192

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2021 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
2:00 182.3 F 2.7  
4:00 197.1 F 7.1  
6:00 212.1 F 12.0  
8:00 224.7 F 7.7  
10:00 237.4 F 12.9  
12:00 238.7 F 11.0  
14:00 238.0 F 9.9  
16:00 251.4 F 16.4  
18:00 262.7 F 22.9  
20:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.1 F 29.0  
2:00 272.2 F 29.5  
3:00 271.3 F 28.7  
4:00 272.1 F 29.2

Min 271.3 F 28.6  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 207.5 F 0.0  
10:00 209.9 F 0.0  
15:00 210.2 F 0.0  
20:00 208.5 F 0.0  
25:00 188.6 F 0.0  
Dry Cycle Aborted

FILLING: 2:00  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 27:56  
TOTAL CYCLE: 00:57:29

07/12/2021 10:31 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01