
Midmark M11 - v1.0.5
Total Cycles: 4231

Name: RITTER2
Sterilizer ID: V2313228
12/19/2024 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.0 F 0.0
2:00 116.0 F 0.3
4:00 132.3 F 2.8
6:00 166.3 F 6.4
8:00 184.8 F 11.1
10:00 210.2 F 4.6
12:00 223.9 F 9.2
14:00 231.8 F 10.7
16:00 237.8 F 10.9
18:00 235.5 F 8.9
20:00 249.6 F 15.4
22:00 261.0 F 22.0
24:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 271.8 F 29.4
2:00 272.5 F 29.3
3:00 271.2 F 28.8
4:00 272.5 F 29.8

Min 271.2 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 209.1 F 0.0
10:00 199.2 F 0.0
15:00 204.0 F 0.0
20:00 212.3 F 0.0
25:00 225.2 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 24:29
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:07
CYCLE COMPLETE

12/19/2024 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____