
Midmark M11 - v1.0.0
Total Cycles: 1641

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2022 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 172.9 F 2.3
4:00 179.6 F 5.8
6:00 200.4 F 10.8
8:00 218.5 F 4.8
10:00 232.8 F 10.3
12:00 232.9 F 7.6
14:00 244.6 F 13.2
16:00 243.2 F 11.7
18:00 256.3 F 18.5
20:00 267.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 28.9
2:00 271.5 F 28.3
3:00 272.6 F 28.9
4:00 271.5 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 231.4 F 0.0
10:00 230.3 F 0.0
15:00 232.2 F 0.0
20:00 224.7 F 0.0
25:00 221.0 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

02/16/2022 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____