

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2493

Name: RITTER 1  
Sterilizer ID: V2178198  
10/26/2022 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
2:00 190.8 F 3.9  
4:00 188.8 F 8.7  
6:00 219.7 F 9.4  
8:00 228.8 F 8.7  
10:00 239.2 F 11.0  
12:00 240.1 F 11.5  
14:00 238.3 F 9.9  
16:00 252.0 F 16.5  
18:00 263.7 F 23.3  
20:00 271.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.7 F 29.5  
2:00 271.9 F 29.1  
3:00 272.8 F 29.6  
4:00 271.9 F 28.9

Min 271.8 F 28.9  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 232.3 F 0.0  
10:00 228.5 F 0.0  
15:00 231.5 F 0.0  
20:00 227.0 F 0.0  
25:00 229.7 F 0.0  
30:00 230.9 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:53  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:29  
CYCLE COMPLETE

10/26/2022 01:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_