

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4421

Name: RITTER2  
Sterilizer ID: V2313228  
03/11/2025 03:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.7 F 0.0  
2:00 98.3 F 0.5  
4:00 106.4 F 2.1  
6:00 127.2 F 4.7  
8:00 161.8 F 7.8  
10:00 172.5 F 11.6  
12:00 207.4 F 4.1  
14:00 211.4 F 7.6  
16:00 223.8 F 11.2  
18:00 229.2 F 7.3  
20:00 238.9 F 12.1  
22:00 237.1 F 10.0  
24:00 248.9 F 15.5  
26:00 259.0 F 21.3  
28:00 268.5 F 27.5  
30:00 272.2 F 30.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.4  
1:00 272.1 F 30.5  
2:00 271.8 F 30.7  
3:00 272.0 F 30.4  
4:00 271.8 F 30.7

Min 271.4 F 30.3  
Max 272.5 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.2 F 0.0  
5:00 222.3 F 0.0  
10:00 223.2 F 0.0  
15:00 227.5 F 0.0  
20:00 218.0 F 0.0  
25:00 192.4 F 0.0  
30:00 183.7 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 30:02  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:08:58  
CYCLE COMPLETE

03/11/2025 05:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_