

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1081

Name: RITTER 1  
Sterilizer ID: V2178198  
08/09/2021 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
2:00 194.3 F 3.5  
4:00 195.0 F 8.9  
6:00 224.7 F 7.4  
8:00 234.7 F 10.2  
10:00 231.7 F 7.1  
12:00 243.4 F 12.4  
14:00 243.7 F 12.1  
16:00 256.5 F 18.8  
18:00 267.6 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.9  
1:00 272.5 F 29.2  
2:00 272.1 F 28.8  
3:00 272.1 F 28.9  
4:00 272.3 F 29.3  
  
Min 271.9 F 28.8  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 231.1 F 0.0  
10:00 223.2 F 0.0  
15:00 223.6 F 0.0  
20:00 206.2 F 0.0  
Dry Cycle Aborted

FILLING: 1:53  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 20:15  
TOTAL CYCLE: 00:49:00

08/09/2021 03:12 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01