
Midmark M11 - v1.0.5
Total Cycles: 41

Name: RITTER2
Sterilizer ID: V2313228
05/20/2021 03:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 164.5 F 2.8
4:00 172.5 F 6.2
6:00 193.9 F 10.3
8:00 213.2 F 6.3
10:00 221.9 F 8.3
12:00 233.4 F 13.2
14:00 238.2 F 11.0
16:00 238.7 F 10.1
18:00 250.9 F 16.3
20:00 262.5 F 22.8
22:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.9
2:00 272.6 F 29.4
3:00 271.5 F 28.9
4:00 271.6 F 29.5

Min 271.3 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 215.0 F 0.0
10:00 219.6 F 0.0
15:00 214.3 F 0.0
20:00 214.3 F 0.0
25:00 211.7 F 0.0
30:00 209.3 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

05/20/2021 04:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____