
Midmark M11 - v1.0.5
Total Cycles: 2393

Name: RITTER2
Sterilizer ID: V2313228
05/19/2023 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.4 F 0.0
2:00 79.8 F 0.0
4:00 89.9 F 0.0
6:00 109.8 F 0.0
8:00 152.8 F 0.3
10:00 185.2 F 0.5
12:00 198.8 F 4.0
14:00 210.0 F 7.3
16:00 220.0 F 10.9
18:00 227.1 F 7.2
20:00 235.3 F 10.3
22:00 236.0 F 9.0
24:00 244.7 F 13.1
26:00 242.7 F 11.8
28:00 254.6 F 17.9
30:00 265.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.3 F 29.4
2:00 272.0 F 28.7
3:00 272.2 F 29.3
4:00 272.3 F 29.0

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 196.5 F 0.0
10:00 213.8 F 0.0
15:00 216.3 F 0.0
20:00 212.7 F 0.0
25:00 213.0 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:52
CYCLE COMPLETE

05/19/2023 08:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____