
Midmark M11 - v1.0.0
Total Cycles: 5322

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2025 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 71.7 F 0.3
4:00 91.6 F 1.9
6:00 131.6 F 4.4
8:00 155.2 F 7.1
10:00 173.5 F 10.5
12:00 185.4 F 5.1
14:00 205.0 F 4.2
16:00 216.7 F 7.6
18:00 227.6 F 11.3
20:00 225.9 F 5.9
22:00 238.5 F 11.1
24:00 234.3 F 8.2
26:00 248.2 F 14.5
28:00 259.4 F 20.7
30:00 269.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.9 F 29.2
2:00 271.7 F 28.4
3:00 272.9 F 29.3
4:00 272.1 F 28.5

Min 271.6 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 188.4 F 0.0
10:00 199.7 F 0.0
15:00 204.5 F 0.0
20:00 205.8 F 0.0
25:00 211.8 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:03
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:14
CYCLE COMPLETE

02/07/2025 10:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____