
Midmark M11 - v1.0.0
Total Cycles: 4747

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2024 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 76.9 F 0.1
4:00 88.3 F 0.2
6:00 125.4 F 1.7
8:00 150.0 F 4.3
10:00 170.6 F 7.6
12:00 188.6 F 11.2
14:00 203.0 F 3.2
16:00 215.0 F 5.7
18:00 225.3 F 9.2
20:00 235.2 F 13.1
22:00 234.5 F 9.0
24:00 240.8 F 10.8
26:00 244.1 F 12.3
28:00 255.7 F 18.3
30:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.4 F 28.1
3:00 272.6 F 28.9
4:00 271.6 F 28.1

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.8 F 0.0
5:00 206.2 F 0.0
10:00 210.5 F 0.0
15:00 219.0 F 0.0
20:00 232.1 F 0.0
25:00 224.6 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:46
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:53
CYCLE COMPLETE

07/19/2024 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____