

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2848

Name: RITTER2  
Sterilizer ID: V2313228  
09/07/2023 05:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
2:00 181.5 F 3.4  
4:00 203.4 F 8.9  
6:00 217.3 F 8.4  
8:00 230.7 F 9.9  
10:00 233.3 F 8.6  
12:00 241.7 F 11.7  
14:00 246.2 F 13.8  
16:00 259.1 F 20.9  
18:00 270.4 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.6 F 29.3  
2:00 272.4 F 29.8  
3:00 271.9 F 29.0  
4:00 272.7 F 29.8  
  
Min 271.5 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 206.0 F 0.0  
10:00 217.7 F 0.0  
15:00 217.7 F 0.0  
20:00 213.3 F 0.0  
25:00 220.5 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 2:19  
HEATING: 18:46  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 00:58:28  
CYCLE COMPLETE

09/07/2023 06:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_