
Midmark M11 - v1.0.0
Total Cycles: 1688

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2022 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.6 F 0.0
2:00 150.5 F 1.0
4:00 167.2 F 3.3
6:00 196.2 F 7.1
8:00 225.8 F 11.8
10:00 224.0 F 5.5
12:00 236.7 F 10.5
14:00 234.1 F 7.7
16:00 245.8 F 13.3
18:00 243.7 F 12.0
20:00 256.8 F 18.8
22:00 268.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.4 F 28.6
2:00 272.8 F 29.0
3:00 271.6 F 28.4
4:00 272.5 F 28.8

Min 271.6 F 28.3
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 224.9 F 0.0
10:00 201.1 F 0.0
15:00 195.2 F 0.0
20:00 193.3 F 0.0
25:00 194.3 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

03/01/2022 04:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____