
Midmark M11 - v1.0.5
Total Cycles: 2809

Name: RITTER2
Sterilizer ID: V2313228
08/29/2023 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 170.9 F 0.0
4:00 190.3 F 2.3
6:00 209.1 F 6.5
8:00 223.9 F 11.0
10:00 229.2 F 6.9
12:00 238.4 F 12.3
14:00 238.3 F 10.3
16:00 237.4 F 8.7
18:00 250.5 F 15.9
20:00 262.2 F 22.6
22:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.1 F 29.1
2:00 272.1 F 29.5
3:00 271.6 F 28.8
4:00 272.5 F 29.5

Min 271.3 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 216.2 F 0.0
10:00 219.5 F 0.0
15:00 225.7 F 0.0
20:00 226.1 F 0.0
25:00 218.0 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

08/29/2023 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____