

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1076

Name: RITTER2  
Sterilizer ID: V2313228  
05/05/2022 09:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.1 F 0.0  
2:00 170.4 F 2.8  
4:00 196.3 F 7.7  
6:00 217.0 F 13.5  
8:00 229.8 F 9.4  
10:00 235.1 F 9.1  
12:00 245.0 F 13.4  
14:00 244.6 F 12.6  
16:00 257.5 F 19.6  
18:00 268.8 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.7 F 29.3  
2:00 272.9 F 29.7  
3:00 271.7 F 29.2  
4:00 272.7 F 29.4  
  
Min 271.7 F 29.0  
Max 273.1 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 175.0 F 0.0  
10:00 150.5 F 0.0  
15:00 141.0 F 0.0  
20:00 135.4 F 0.0  
25:00 131.7 F 0.0  
30:00 130.4 F 0.0  
  
Drying Complete

FILLING: 2:21  
HEATING: 19:52  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:25  
CYCLE COMPLETE

05/05/2022 10:22 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02