
Midmark M11 - v1.0.5
Total Cycles: 2060

Name: RITTER2
Sterilizer ID: V2313228
02/16/2023 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 79.7 F 0.0
4:00 94.8 F 0.0
6:00 116.8 F 0.0
8:00 162.6 F 0.0
10:00 193.2 F 2.1
12:00 203.1 F 5.0
14:00 214.7 F 8.4
16:00 226.5 F 12.2
18:00 230.0 F 7.7
20:00 240.8 F 12.6
22:00 238.7 F 10.4
24:00 236.1 F 8.5
26:00 249.1 F 15.1
28:00 260.3 F 21.3
30:00 270.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.3 F 29.3
2:00 271.4 F 28.9
3:00 272.4 F 29.3
4:00 271.4 F 28.7

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 190.2 F 0.0
10:00 203.6 F 0.0
15:00 213.1 F 0.0
20:00 222.2 F 0.0
25:00 219.8 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:47
CYCLE COMPLETE

02/16/2023 10:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____