
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002590
14:32 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:54
STERILIZING 3:00
136.4 C 226kPa 3:00
136.1 C 223kPa 3:30
136.4 C 226kPa 4:00
136.3 C 225kPa 4:30
136.4 C 227kPa 5:00
136.4 C 226kPa 5:30
136.3 C 225kPa 6:00
136.4 C 225kPa 6:30
136.4 C 225kPa 6:30
VENTING 6:30
AIR DRYING 7:23
CYCLE COMPLETE 37:23

Digital Signature #
D669DA2CBBD63BBC
