
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002892
8:25 29/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 4:07
136.3 C 221kPa 4:07
136.3 C 222kPa 4:37
136.1 C 221kPa 5:07
136.2 C 222kPa 5:37
136.4 C 224kPa 6:07
136.4 C 220kPa 6:37
136.4 C 222kPa 7:07
136.1 C 217kPa 7:37
136.1 C 217kPa 7:37
VENTING 7:37
CYCLE ABORTED 7:44
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:44

Digital Signature #
4CF9011AAA0021B2
