
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000713
7:09 21/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:05
STERILIZING 3:02
136.4 C 232kPa 3:02
136.4 C 232kPa 3:32
136.4 C 232kPa 4:02
136.4 C 230kPa 4:32
136.3 C 231kPa 5:02
136.4 C 232kPa 5:32
136.1 C 228kPa 6:02
136.4 C 232kPa 6:32
136.4 C 232kPa 6:32
VENTING 6:32
AIR DRYING 7:20
CYCLE COMPLETE 7:35

Digital Signature #
24056C32E4123E8A
