

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4505

Name: RITTER 1  
Sterilizer ID: V2178198  
04/30/2024 10:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.0 F 0.0  
2:00 169.3 F 2.2  
4:00 178.6 F 5.6  
6:00 190.1 F 9.8  
8:00 204.9 F 9.0  
10:00 221.9 F 6.8  
12:00 231.7 F 11.0  
14:00 231.8 F 6.6  
16:00 241.3 F 12.3  
18:00 239.5 F 10.3  
20:00 252.4 F 16.6  
22:00 263.5 F 23.1  
24:00 272.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.0  
2:00 271.5 F 28.3  
3:00 272.8 F 29.2  
4:00 272.0 F 28.6

Min 271.4 F 28.3  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 205.5 F 0.0  
10:00 197.4 F 0.0  
15:00 198.7 F 0.0  
20:00 198.1 F 0.0  
25:00 199.7 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:15  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:02:47  
CYCLE COMPLETE

04/30/2024 11:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_