

-----  
Midmark M11 - v1.0.5  
Total Cycles: 598

Name: RITTER2  
Sterilizer ID: V2313228  
12/01/2021 04:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
2:00 178.6 F 3.4  
4:00 197.5 F 8.6  
6:00 216.6 F 9.2  
8:00 228.2 F 9.8  
10:00 232.7 F 8.5  
12:00 242.7 F 12.2  
14:00 245.4 F 13.4  
16:00 258.4 F 20.5  
18:00 269.9 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.1  
1:00 272.2 F 29.2  
2:00 272.3 F 29.3  
3:00 272.6 F 29.5  
4:00 272.8 F 29.8  
  
Min 271.7 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 243.4 F 0.0  
10:00 221.9 F 0.0  
15:00 222.7 F 0.0  
20:00 218.0 F 0.0  
25:00 215.0 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 2:31  
HEATING: 18:43  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

12/01/2021 05:55 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02