

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4040

Name: RITTER 1  
Sterilizer ID: V2178198  
11/20/2023 11:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 172.5 F 2.7  
4:00 188.4 F 6.5  
6:00 202.4 F 11.5  
8:00 216.7 F 5.0  
10:00 225.3 F 9.7  
12:00 234.0 F 9.4  
14:00 239.8 F 11.3  
16:00 238.2 F 9.5  
18:00 251.1 F 15.7  
20:00 262.2 F 22.1  
22:00 272.2 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 29.1  
2:00 272.2 F 28.6  
3:00 271.2 F 28.2  
4:00 272.1 F 28.9

Min 271.2 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 237.5 F 0.0  
10:00 220.3 F 0.0  
15:00 215.4 F 0.0  
20:00 218.3 F 0.0  
25:00 219.7 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 22:32  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:01:51  
CYCLE COMPLETE

11/20/2023 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_