

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2220

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2022 09:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.4 F 0.0  
2:00 135.0 F 0.9  
4:00 146.6 F 3.8  
6:00 171.9 F 7.7  
8:00 194.2 F 12.0  
10:00 210.9 F 4.3  
12:00 222.0 F 8.5  
14:00 233.1 F 12.8  
16:00 233.9 F 8.9  
18:00 241.4 F 11.3  
20:00 243.8 F 12.2  
22:00 255.6 F 18.4  
24:00 266.4 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.0  
2:00 271.6 F 28.3  
3:00 272.1 F 29.0  
4:00 272.3 F 28.8

Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 222.1 F 0.0  
10:00 228.2 F 0.0  
15:00 230.9 F 0.0  
20:00 225.5 F 0.0  
25:00 228.3 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 25:39  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:03:37  
CYCLE COMPLETE

08/15/2022 10:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_