
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000986
13:01 10/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:13
STERILIZING 3:14
136.0 C 229kPa 3:14
136.2 C 231kPa 3:44
136.2 C 231kPa 4:14
136.4 C 232kPa 4:44
136.4 C 233kPa 5:14
136.4 C 233kPa 5:44
136.4 C 231kPa 6:14
136.4 C 233kPa 6:44
136.4 C 233kPa 6:44
VENTING 6:44
AIR DRYING 7:37
CYCLE COMPLETE 8:26

Digital Signature #
8E16A6EC8C9D29E9
