

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002543  
8:55 28/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:03  
STERILIZING 3:07  
136.4 C 222kPa 3:07  
136.2 C 223kPa 3:37  
136.4 C 223kPa 4:07  
136.2 C 222kPa 4:37  
136.4 C 221kPa 5:07  
136.5 C 223kPa 5:37  
136.2 C 223kPa 6:07  
136.0 C 219kPa 6:37  
136.0 C 219kPa 6:37  
VENTING 6:37  
AIR DRYING 7:29  
CYCLE COMPLETE 30:47  
-----

Digital Signature #  
4E0077F488C1217D  
-----