
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008219
8:30 22/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:54
STERILIZING 3:47
136.1 C 220kPa 3:47
136.2 C 221kPa 4:17
136.0 C 221kPa 4:47
136.2 C 222kPa 5:17
136.4 C 222kPa 5:47
136.4 C 222kPa 6:17
136.4 C 220kPa 6:47
136.4 C 222kPa 7:17
136.4 C 222kPa 7:17
VENTING 7:17
AIR DRYING 8:15
CYCLE COMPLETE 38:15

Digital Signature #
869ACF7E05A9D65B
