

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1327

Name: RITTER 1  
Sterilizer ID: V2178198  
11/01/2021 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 184.4 F 3.9  
4:00 200.6 F 9.8  
6:00 221.3 F 6.2  
8:00 231.2 F 12.3  
10:00 236.1 F 10.0  
12:00 235.2 F 8.1  
14:00 249.8 F 15.4  
16:00 261.9 F 22.4  
18:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.0  
1:00 272.3 F 29.2  
2:00 272.5 F 29.4  
3:00 272.6 F 29.6  
4:00 271.8 F 29.3  
  
Min 271.6 F 28.9  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 232.0 F 0.0  
10:00 219.0 F 0.0  
15:00 217.2 F 0.0  
20:00 215.5 F 0.0  
25:00 214.5 F 0.0  
30:00 214.2 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 18:51  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:44  
CYCLE COMPLETE

11/01/2021 11:30 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01