

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2213

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2023 10:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.5 F 0.0  
2:00 129.5 F 1.3  
4:00 138.7 F 3.9  
6:00 166.7 F 7.8  
8:00 193.7 F 12.9  
10:00 223.6 F 7.2  
12:00 236.7 F 12.3  
14:00 234.2 F 9.2  
16:00 237.9 F 9.7  
18:00 246.7 F 13.9  
20:00 258.7 F 20.6  
22:00 269.8 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.7 F 30.4  
2:00 272.9 F 30.2  
3:00 272.9 F 30.1  
4:00 272.8 F 30.1

Min 270.8 F 28.8  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.7 F 0.0  
5:00 227.4 F 0.0  
10:00 245.8 F 0.0  
15:00 222.3 F 0.0  
20:00 211.5 F 0.0  
25:00 219.2 F 0.0  
30:00 217.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 22:47  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:53  
CYCLE COMPLETE

04/04/2023 11:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_