

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1845

Name: RITTER 1  
Sterilizer ID: V2178198  
04/20/2022 08:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.9 F 0.0  
2:00 157.8 F 1.4  
4:00 172.1 F 5.2  
6:00 198.3 F 10.0  
8:00 219.8 F 4.8  
10:00 233.6 F 9.9  
12:00 233.1 F 7.5  
14:00 243.6 F 12.6  
16:00 241.6 F 10.9  
18:00 255.1 F 17.8  
20:00 267.0 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.5 F 28.8  
2:00 271.3 F 28.0  
3:00 272.7 F 28.8  
4:00 271.6 F 28.1  
  
Min 271.2 F 28.0  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
5:00 237.5 F 0.0  
10:00 220.9 F 0.0  
15:00 221.0 F 0.0  
20:00 222.6 F 0.0  
25:00 239.0 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 00:59:25  
CYCLE COMPLETE

04/20/2022 09:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_