
Midmark M11 - v1.0.5
Total Cycles: 2063

Name: RITTER2
Sterilizer ID: V2313228
02/16/2023 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.4 F 0.0
2:00 130.8 F 0.0
4:00 158.7 F 1.0
6:00 182.1 F 4.5
8:00 205.1 F 8.6
10:00 224.0 F 13.4
12:00 228.9 F 9.0
14:00 236.7 F 10.6
16:00 241.8 F 12.2
18:00 240.6 F 11.0
20:00 253.6 F 17.6
22:00 264.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.5
2:00 271.2 F 28.8
3:00 272.5 F 29.7
4:00 272.0 F 29.1

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 208.6 F 0.0
10:00 218.7 F 0.0
15:00 223.2 F 0.0
20:00 226.5 F 0.0
25:00 223.9 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:51
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

02/16/2023 02:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____