
Midmark M11 - v1.0.0
Total Cycles: 3621

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2023 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	69.0 F	0.0
2:00	77.3 F	0.0
4:00	100.0 F	0.0
6:00	111.1 F	0.0
8:00	140.6 F	0.1
10:00	169.0 F	1.8
12:00	188.4 F	4.5
14:00	211.0 F	7.5
16:00	213.8 F	10.7
18:00	228.6 F	8.3
20:00	229.8 F	8.1
22:00	237.4 F	12.2
24:00	235.2 F	8.6
26:00	244.5 F	12.3
28:00	243.4 F	11.9
30:00	254.8 F	17.7
32:00	264.9 F	23.8
34:00	271.2 F	28.1

STERILIZING

mm:ss	Degrees	PSI
0:00	271.2 F	28.2
1:00	272.5 F	28.7
2:00	271.3 F	28.3
3:00	272.5 F	28.7
4:00	271.2 F	28.1

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	192.2 F	0.0
	5:00	195.8 F	0.0
	10:00	197.1 F	0.0
	15:00	208.5 F	0.0
	20:00	220.3 F	0.0
Drying Complete	25:00	228.4 F	0.0
	30:00	228.7 F	0.0

FILLING: 0:47
HEATING: 34:01
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:12:06
CYCLE COMPLETE

08/08/2023 11:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____