
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000175
14:08 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:15
PRESSURIZING 4:15
STERILIZING 5:21
136.0 C 221kPa 5:21
136.4 C 226kPa 5:51
136.4 C 225kPa 6:21
136.4 C 225kPa 6:51
136.4 C 224kPa 7:21
136.4 C 226kPa 7:51
136.3 C 225kPa 8:21
136.3 C 224kPa 8:51
136.3 C 224kPa 8:51
VENTING 8:51
AIR DRYING 9:51
CYCLE COMPLETE 24:51

Digital Signature #
95F77909F6FFB2B5
