

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000417  
8:38 27/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:50  
STERILIZING 4:03  
136.0 C 228kPa 4:03  
136.4 C 233kPa 4:33  
136.2 C 231kPa 5:03  
136.4 C 233kPa 5:33  
136.3 C 232kPa 6:03  
136.4 C 232kPa 6:33  
136.3 C 233kPa 7:03  
136.2 C 232kPa 7:33  
136.2 C 232kPa 7:33  
VENTING 7:33  
AIR DRYING 8:26  
CYCLE COMPLETE 8:34

-----  
Digital Signature #  
4D2C0022AF0D1B64  
-----