

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4749

Name: RITTER 1  
Sterilizer ID: V2178198  
07/19/2024 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.4 F 0.0  
2:00 122.3 F 1.4  
4:00 136.0 F 3.9  
6:00 163.2 F 6.9  
8:00 186.7 F 10.3  
10:00 209.7 F 8.2  
12:00 214.8 F 6.5  
14:00 226.2 F 10.1  
16:00 233.5 F 8.9  
18:00 236.7 F 10.4  
20:00 239.8 F 8.7  
22:00 245.8 F 13.3  
24:00 256.5 F 18.9  
26:00 266.2 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.1 F 29.2  
2:00 272.7 F 29.4  
3:00 272.4 F 29.1  
4:00 271.9 F 28.9

Min 271.2 F 28.4  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 200.1 F 0.0  
10:00 204.0 F 0.0  
15:00 207.0 F 0.0  
20:00 208.6 F 0.0  
25:00 210.6 F 0.0  
30:00 210.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 27:56  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:06:44  
CYCLE COMPLETE

07/19/2024 11:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_