
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001585
10:37 13/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:12
STERILIZING 3:16
136.4 C 225kPa 3:16
136.2 C 224kPa 3:46
136.0 C 219kPa 4:16
136.1 C 223kPa 4:46
136.3 C 225kPa 5:16
136.4 C 225kPa 5:46
136.4 C 226kPa 6:16
136.5 C 226kPa 6:46
136.5 C 226kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CYCLE COMPLETE 7:42

Digital Signature #
0FE342D40ACE4457
