
Midmark M11 - v1.0.0
Total Cycles: 2211

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2022 05:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.7 F 0.0
2:00 203.5 F 4.8
4:00 203.7 F 9.9
6:00 224.3 F 6.0
8:00 232.0 F 10.8
10:00 233.4 F 7.9
12:00 242.5 F 11.3
14:00 245.8 F 13.3
16:00 258.6 F 20.1
18:00 269.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.0
2:00 271.4 F 28.4
3:00 272.5 F 29.2
4:00 271.8 F 28.5

Min 271.3 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 232.0 F 0.0
10:00 239.4 F 0.0
15:00 238.2 F 0.0
20:00 227.2 F 0.0
25:00 225.3 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

08/08/2022 06:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____