
Midmark M11 - v1.0.0
Total Cycles: 3121

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.3 F 0.0
2:00 194.6 F 0.0
4:00 204.8 F 5.2
6:00 220.6 F 10.4
8:00 227.1 F 6.6
10:00 238.8 F 12.1
12:00 238.2 F 10.2
14:00 235.8 F 8.5
16:00 251.0 F 15.8
18:00 263.0 F 22.8
20:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.4 F 28.9
2:00 272.0 F 28.9
3:00 272.3 F 28.8
4:00 271.9 F 28.9

Min 271.6 F 28.4
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 228.9 F 0.0
10:00 236.8 F 0.0
15:00 230.1 F 0.0
20:00 226.6 F 0.0
25:00 225.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

04/18/2023 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____