

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1370

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2021 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
2:00 185.0 F 3.3  
4:00 200.1 F 8.0  
6:00 217.3 F 10.3  
8:00 232.7 F 9.2  
10:00 238.1 F 9.6  
12:00 244.2 F 12.9  
14:00 243.1 F 11.8  
16:00 256.6 F 18.8  
18:00 268.2 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.0 F 28.5  
2:00 272.3 F 29.0  
3:00 271.3 F 28.5  
4:00 272.1 F 28.7  
  
Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
5:00 229.8 F 0.0  
10:00 227.4 F 0.0  
15:00 230.6 F 0.0  
20:00 206.6 F 0.0  
25:00 195.0 F 0.0  
30:00 190.8 F 0.0  
Drying Complete

FILLING: 2:09  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

11/15/2021 01:29 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02