
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004424
12:57 22/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:18
STERILIZING 3:21
136.0 C 220kPa 3:21
136.0 C 222kPa 3:51
136.3 C 224kPa 4:21
136.4 C 225kPa 4:51
136.4 C 222kPa 5:21
136.2 C 224kPa 5:51
136.4 C 224kPa 6:21
136.3 C 223kPa 6:51
136.3 C 223kPa 6:51
VENTING 6:51
AIR DRYING 7:40
CYCLE COMPLETE 7:49

Digital Signature #
7BC6B6C82F23983B
