

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001303  
14:54 19/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:47  
PRESSURIZING 2:59  
STERILIZING 4:12  
136.0 C 220kPa 4:12  
136.3 C 224kPa 4:42  
136.2 C 222kPa 5:12  
136.1 C 222kPa 5:42  
136.2 C 223kPa 6:12  
136.4 C 224kPa 6:42  
136.3 C 224kPa 7:12  
135.9 C 221kPa 7:42  
135.9 C 221kPa 7:42  
VENTING 7:42  
AIR DRYING 8:33  
CYCLE COMPLETE 9:33  
-----

Digital Signature #  
65FD7FBE81C6F72A  
-----