
Midmark M11 - v1.0.5
Total Cycles: 2099

Name: RITTER2
Sterilizer ID: V2313228
02/28/2023 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.2 F 0.0
2:00 162.2 F 1.9
4:00 167.5 F 5.1
6:00 189.6 F 9.7
8:00 215.8 F 7.0
10:00 225.6 F 9.9
12:00 230.6 F 8.0
14:00 242.3 F 13.3
16:00 243.0 F 12.4
18:00 255.6 F 19.3
20:00 267.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.9 F 30.0
2:00 271.5 F 29.3
3:00 272.2 F 30.2
4:00 272.4 F 29.8

Min 271.2 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 217.0 F 0.0
10:00 232.0 F 0.0
15:00 229.6 F 0.0
20:00 224.5 F 0.0
25:00 226.4 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

02/28/2023 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____