
Midmark M11 - v1.0.5
Total Cycles: 2535

Name: RITTER2
Sterilizer ID: V2313228
06/20/2023 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 77.3 F 0.0
4:00 108.5 F 0.0
6:00 110.9 F 0.0
8:00 145.5 F 0.4
10:00 170.8 F 3.6
12:00 187.6 F 7.1
14:00 205.7 F 10.8
16:00 218.8 F 6.5
18:00 226.8 F 8.5
20:00 236.8 F 12.7
22:00 235.8 F 9.6
24:00 239.1 F 9.9
26:00 245.9 F 13.4
28:00 257.1 F 19.4
30:00 267.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.6
2:00 271.7 F 28.7
3:00 272.1 F 29.5
4:00 272.3 F 29.2

Min 271.4 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 209.4 F 0.0
10:00 215.2 F 0.0
15:00 210.4 F 0.0
20:00 218.1 F 0.0
25:00 214.5 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:35
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

06/20/2023 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____