
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003036
10:00 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:40
136.4 C 220kPa 4:40
136.4 C 225kPa 5:10
136.4 C 220kPa 5:40
136.4 C 221kPa 6:10
136.4 C 221kPa 6:40
136.2 C 221kPa 7:10
136.4 C 223kPa 7:40
136.3 C 218kPa 8:10
136.3 C 218kPa 8:10
VENTING 8:10
AIR DRYING 9:06
CYCLE COMPLETE 9:12

Digital Signature #
7D8537FC18D2B125
