

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001699  
9:12 08/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:24  
PRESSURIZING 3:24  
STERILIZING 4:26  
136.0 C 220kPa 4:26  
136.1 C 222kPa 4:56  
136.4 C 223kPa 5:26  
136.3 C 223kPa 5:56  
136.4 C 223kPa 6:26  
136.4 C 223kPa 6:56  
136.4 C 223kPa 7:26  
136.1 C 222kPa 7:56  
136.1 C 222kPa 7:56  
VENTING 7:56  
AIR DRYING 8:55  
CYCLE COMPLETE 9:09

-----  
Digital Signature #  
DD352060F7F73B6C  
-----