
Midmark M11 - v1.0.0
Total Cycles: 811

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 06:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 79.4 F 0.0
4:00 81.9 F 0.0
6:00 111.6 F 0.0
8:00 143.2 F 0.2
10:00 174.1 F 0.2
12:00 191.4 F 1.6
14:00 203.9 F 4.1
16:00 212.8 F 6.8
18:00 224.3 F 10.3
20:00 231.2 F 9.6
22:00 234.2 F 9.5
24:00 240.8 F 10.6
26:00 243.5 F 12.2
28:00 241.6 F 10.8
30:00 253.6 F 16.9
32:00 264.2 F 23.2
34:00 272.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.8
2:00 272.1 F 28.2
3:00 271.7 F 28.3
4:00 272.7 F 28.8

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 200.2 F 0.0
10:00 203.1 F 0.0
15:00 213.0 F 0.0
20:00 220.0 F 0.0
25:00 224.2 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 34:10
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:12:35
CYCLE COMPLETE

04/27/2021 07:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____