

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2247

Name: RITTER2  
Sterilizer ID: V2313228  
04/13/2023 12:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
2:00 180.9 F 3.8  
4:00 193.3 F 8.5  
6:00 211.7 F 9.8  
8:00 227.7 F 9.4  
10:00 233.6 F 9.2  
12:00 243.2 F 13.3  
14:00 243.3 F 12.6  
16:00 256.2 F 19.4  
18:00 267.8 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.5 F 29.8  
2:00 271.5 F 29.3  
3:00 272.6 F 29.9  
4:00 271.8 F 29.2  
  
Min 271.5 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 232.4 F 0.0  
10:00 242.9 F 0.0  
15:00 233.3 F 0.0  
20:00 223.6 F 0.0  
25:00 223.1 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 19:18  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

04/13/2023 01:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_